



Facility Permit Inspection

Name of Licensee: _____

License #: _____

Date: _____

Facility Address: _____

Name of Inspector: _____

Oxygen Delivery Systems- Den 304.07 (a) (1)			
Requirement	Yes	No	
Portable O ₂ tank			
Bag valve mask with connector tubing for O ₂ tank			
Nasal cannula or oxygen mask with connector tubing for O ₂ tank			
Suction Equipment- Den 304.07 (a)(2)			
Requirement	Yes	No	
Capable of suctioning the throat in all operatories and recovery rooms			
Yankauer suction tip with connector for suction unit			
Transportation Equipment- Den 304.07 (a)(3)			
Requirement	Yes	No	
Wheelchair			
Vital Signs Monitoring Equipment- Den 304.07 (a)(4)			
Requirement	NA	Yes	No
Continuous pulse oximeter			
Blood pressure monitor with cuffs for adult patients*			
Blood pressure monitor with cuffs for patients under 13*			
Defibrillator Equipment- Den 304.07 (a) (5)			
Requirement	NA	Yes	No
Manual or automated external defibrillator (AED)			
Unexpired batteries for defibrillator or AED, if battery powered			
Log maintained documenting monthly inspection for defibrillator or AED functionality			
Unexpired adult defibrillator pads that connect to the facility's defibrillator or AED			
Unexpired pediatric defibrillator pads that connect to the facility's defibrillator or AED for patients under 13			

*Ability to Monitor Blood Pressure



STATE OF NEW HAMPSHIRE
OFFICE OF PROFESSIONAL LICENSURE AND CERTIFICATION
BOARD OF DENTAL EXAMINERS
7 EAGLE SQUARE, CONCORD, NH 03301-4980
Telephone: 603-271-2152
TDD Access: Relay NH 1-800-735-2964
www.oplc.nh.gov

Auxiliary Light Source- Den 304.07 (a) (6)			
Requirement	Yes	No	
Auxiliary light source capable of use during power failure			
Rigid CPR Board- Den 304.07 (a) (7)			
Requirement	Yes	No	
Rigid CPR Board			
Staffing Requirements- Den 304.07 (a) (8)			
Requirement	NA	Yes	No
Operating Dentist PALS certification, if not the qualified provider and if patients under age 13 are treated at facility			
Dental Assistant with current BLS Certification on staff for procedure			
Additional Personnel available to call 911			
Appropriate staffing on site, as described in Den 304.05			
Oxygen & Gas Delivery Systems- Den 304.07 (b) (1)			
Requirement	Yes	No	
Oxygen fail-safe valve on anesthesia machine or N ₂ O flow meter if administered			
Size-dissimilar fittings, varied hose sizes, and colored tubing and hoses for oxygen, nitrous oxide, and medical air if used			
Suction Systems- Den 304.07(b)(2)			
Requirement	Yes	No	
Pulmonary suction catheter with connector tubing able to be connected to suction unit or vacuum system.			
Portable suction device able to be used during power failure			
Airway Devices- Den 304.07 (b)(3)			
Requirement	NA	Yes	No
Endotracheal tubes (ETT) with stylets for adult patients			
Endotracheal tubes with stylets for patients under 13 years of age			
Syringe for ETT cuff inflation			
Ambu bag and connector tubing to enable ETT to be connected to oxygen tank and deliver positive pressure			
Video laryngoscope with blades of various sizes for adult patients			
Video laryngoscope with blades of various sizes for patients under 13 years of age			
Extra batteries for laryngoscope			
Extra bulbs for laryngoscope, if applicable			
Magill forceps			
Tongue grasping forceps			
Oral airways for adult patients			
Oral airways for patients under 13 years of age if treated at the facility			
Nasopharyngeal airways for adult patients			
Nasopharyngeal airways patients under 13 years of age if treated at the facility			
Supraglottic airways (such as Laryngeal Mask Airways or King Airways) for adult patients			



STATE OF NEW HAMPSHIRE
OFFICE OF PROFESSIONAL LICENSURE AND CERTIFICATION
BOARD OF DENTAL EXAMINERS
7 EAGLE SQUARE, CONCORD, NH 03301-4980
Telephone: 603-271-2152
TDD Access: Relay NH 1-800-735-2964
www.oplc.nh.gov

Supraglottic airways (such as Laryngeal Mask Airways or King Airways) for patients under 13 years of age			
Emergency cricothyrotomy kit with connector tubing able to be connected to bag valve mask and oxygen tank			
Vital Signs Monitoring Equipment in Operatories- Den 304.07 (b) (4)			
Requirement		Yes	No
Continuous electrocardiogram			
Capnography			
Stethoscope			
Accessory Supplies And Emergency Medications- Den 304.07 (b) (5)			
Requirement		Yes	No
Disposable angiocatheters, tourniquets, and tape for IV catheter placement			
IV tubing and IV fluids			
Disposable needles and syringes for medication draw up and delivery			
Epinephrine 1:10,000 for cardiac arrest			
Epinephrine 1:1,000 for allergic reaction			
Direct or indirect acting pressor for hypotension			
Antiarrhythmic for cardiac arrest			
Adenosine for tachycardia			
Naloxone for opioid reversal			
Flumazenil for benzodiazepine reversal			
Antihypertensive for acute hypertension			
Steroid injection for allergic reaction			
Dextrose injection for hypoglycemic reversal			
Atropine for bradycardia			
Diphenhydramine injection for allergic reaction			
Neuromuscular blocker with rapid onset for laryngospasm			
Albuterol inhaler for bronchospasm			
Benzodiazepine injection for seizure			
Immediate-release non-enteric coated aspirin for acute coronary syndrome			
Sublingual nitroglycerin for acute coronary syndrome			
ONLY NEEDED IF PROVIDER WILL BE ADMINISTERING INHALED VOLATILE ANESTHETICS, OR SUCCINYLCHOLINE IF ON A ROUTINE BASIS - Den 304.07 (c)			
Requirement	NA	Yes	No
Oxygen and gas delivery system also with anesthetic gas analyzer if inhaled volatile anesthetics are used			
Dantrolene injection for malignant hyperthermia			
Sterile water for dantrolene diluent			

For Inspector's Use Only



STATE OF NEW HAMPSHIRE
OFFICE OF PROFESSIONAL LICENSURE AND CERTIFICATION
BOARD OF DENTAL EXAMINERS
7 EAGLE SQUARE, CONCORD, NH 03301-4980
Telephone: 603-271-2152
TDD Access: Relay NH 1-800-735-2964
www.oplc.nh.gov

Deficiencies:

Correction Plan:

Signature: _____

Date: _____

DRAFT